



Lauren Studdard <lstuddard@tsbde.texas.gov>

Lead Apron use

Lori Fulk <fulk1121@gmail.com>
To: stakeholders@tsbde.texas.gov

Wed, Aug 28, 2024 at 10:35 AM

I agree that the Board Rules should be amended. What I've seen most is the request of using a thyroid collar during panorex xrays and all that happens is that it obscures the view of the lower incisors as well as the mandibular symphysis. Thus, a new xray has to be taken anyway resulting in excess radiation.

Sincerely,

Lori Fulk (license #22556)