



Lauren Studdard <lstuddard@tsbde.texas.gov>

Lead shielding for xrays

1 message

Dr Mark Stephenson DDS <drmark@hospitalsedation.com>
To: stakeholders@tsbde.texas.gov

Wed, Aug 28, 2024 at 4:32 AM

When the handheld x-ray units (Nomad) first became available I was hesitant about any x-ray exposure that could happen to me or my assistant. I ordered and wore x-ray exposure badges for a year and sent them in for evaluation. Results were negative for any exposure. That was 15 years ago. I agree with the ADA stance on their new policies regarding no need for lead aprons for dental x-rays if they are digital. I trust that many research results were evaluated by the ADA and this decision was taken very seriously.

--

Dr. Mark Stephenson DDS
500 MEDICAL CENTER BLVD
SUITE 110
CONROE, TX 77304
office (936) 760-2200

THIS EMAIL ADDRESS USES GOOGLE G SUITE ENCRYPTION WHICH IS HIPPA COMPLIANT
fax (936) 760-2226
www.hospitalsedation.com